EU-Declaration of Conformity

Document No. 320CE24121

We, RIKEN KEIKI Co., Ltd. 2-7-6, Azusawa, Itabashi-ku, Tokyo, 174-8744, Japan
declare under our sole responsibility that the following product conforms to
all the relevant provisions.

Product Name Portable Gas Monitor
Model OX-04, OX-04G, HS-04, CO-04, CX-04, SC-04

Council Directives	Applicable Standards
EMC Directive （2014/30/EU）	EN 50270:2015
ATEX Directive （2014/34/EU）	EN IEC 60079-0:2018 EN 60079-11:2012
BATTERY Regulation （(EU)2023/1542）	-
RoHS Directive （2011/65/EU[1]）	EN IEC 63000:2018

[1]Including substances added by Commission Delegated Directive (EU) 2015/863

EU-Type examination Certificate No. DEKRA 19ATEX0097

Notified Body for ATEX
DEKRA Certification B.V. (NB 0344)
Meander 1051, 6825 MJ Arnhem
P.O. Box 5185, 6802 ED Arnhem
The Netherlands

Auditing Organization for ATEX
DEKRA Certification B.V. (NB 0344)
Meander 1051, 6825 MJ Arnhem
P.O. Box 5185, 6802 ED Arnhem
The Netherlands

The marking of the product shall include the following:

Ex II 1 G Ex ia IIC T4...T3 Ga -40℃ ≤ Ta ≤ +60℃

Alternative Marking: T4: when equiped with primary batteries
T3: when equiped with secondary batteries

Place: Tokyo, Japan

Date: Nov. 08, 2024

Takakura Toshiyuki
General manager
Quality Control Center